

The Neurogram - A quantification of real-life hearing impairments using electrophysiology

Fabian Schmidt^{1,2}, Lisa Reisinger^{1,2}, Patrick Neff^{1,3}, Ronny Hannemann⁴, & Nathan Weisz^{1,2}

1. Centre for Cognitive Neuroscience, University of Salzburg, 5020 Salzburg, Austria
2. Department of Psychology, University of Salzburg, 5020 Salzburg, Austria
3. Institute of Bioengineering, Center for Neuroprosthetics, École Polytechnique Fédérale de Lausanne, Switzerland
4. Audiological Research Unit, WSAudiology - Sivantos GmbH, 91058 Erlangen, Germany

1 Motivation

Pure-Tone Audiometry (PTA):

- Assesses hearing thresholds in dB using artificial pure tones (log-spaced between 125 and 8000Hz)
- Information about hearing thresholds is obtained via subjective feedback


Problem:

- Artificial pure-tones do not reflect real-life listening situations (e.g. cocktail party)
- Supra-threshold hearing loss (i.e. hidden hearing loss) is not captured using PTA
- Subjective feedback problematic for babies or old people suffering from dementia

2 Material & Methods


- N=43 subjects
- Age = 43.49
- Online Hearing Assessment
- Stimulus material: radio play (~20 min)
- 306 channel MEG system

audio material (radio play)




Analysis Procedure

1) Acoustic Feature Extraction




2) Encoding Model



3) Audiogram | EEG/MEG Prediction




4) Decoding Model




3

Spectrogram reconstruction accuracy at selected channels can be related to individual hearing levels



Magnetometers


Gradiometers



- Channels with spectrogram encoding accuracies that correlate with pure-tone audiometry scores are selected for reconstruction
- Spectrogram reconstruction accuracy at selected channels is negatively related to pure-tone audiometry scores

5


Spatial hearing is the best predictor for a strong relationship between Neurogram and Audiogram



- High variance in goodness of fit (R^2) between Neurogram and Audiogram across subjects
- Using several predictors (SSQ-Scores, Age etc.) to explain the goodness of fit we find that:
 - Goodness of Neurogram/Audiogram fit is explained best by subjective reports of spatial hearing abilities

4

Neurogram scores are stronger related to subjective reports of hearing impairment than audiogram scores



- Subjective reports of hearing impairment were assessed via the speech, spatial and qualities of hearing scale (SSQ) and correlated with neurogram average and pure-tone average scores
- Posterior distributions for the correlation coefficients of neurogram average and pure-tone average scores with the SSQ scales are compared
- Neurogram average scores are stronger related to SSQ scores than audiogram average scores